
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001223
8:39 22/02/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:15
PRESSURIZING 4:16
STERILIZING 5:30
136.4 C 222kPa 5:30
136.3 C 221kPa 6:00
136.4 C 224kPa 6:30
136.0 C 220kPa 7:00
136.2 C 218kPa 7:30
136.4 C 223kPa 8:00
136.0 C 221kPa 8:30
136.1 C 222kPa 9:00
136.1 C 222kPa 9:00
VENTING 9:00
AIR DRYING 9:53
CYCLE COMPLETE 9:57

Digital Signature #
C171BC8A8EEC45F3
