
Midmark M11 - v1.0.5
Total Cycles: 1875

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 169.9 F 3.8
4:00 191.2 F 8.1
6:00 209.2 F 13.3
8:00 223.5 F 7.6
10:00 234.8 F 12.8
12:00 235.9 F 10.2
14:00 235.6 F 8.5
16:00 249.2 F 15.4
18:00 261.3 F 22.1
20:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.2 F 29.8
2:00 272.3 F 29.4
3:00 271.7 F 28.9
4:00 271.2 F 28.9

Min 271.2 F 28.7
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 222.9 F 0.0
10:00 235.6 F 0.0
15:00 239.9 F 0.0
20:00 229.2 F 0.0
25:00 225.0 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

12/21/2022 05:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____