
Midmark M11 - v1.0.5
Total Cycles: 1872

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.8 F 0.0
2:00 155.4 F 2.9
4:00 169.2 F 6.6
6:00 195.7 F 11.6
8:00 219.3 F 6.0
10:00 231.1 F 11.5
12:00 232.2 F 8.4
14:00 240.2 F 11.2
16:00 244.1 F 13.0
18:00 256.9 F 19.7
20:00 268.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 30.1
2:00 272.3 F 29.7
3:00 271.5 F 29.1
4:00 271.3 F 29.4

Min 271.1 F 29.0
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 217.5 F 0.0
10:00 231.0 F 0.0
15:00 228.0 F 0.0
20:00 227.3 F 0.0
25:00 228.6 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

12/21/2022 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____