
Midmark M11 - v1.0.5
Total Cycles: 4321

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 05:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.4 F 0.0
2:00 137.1 F 2.3
4:00 161.8 F 6.2
6:00 189.7 F 11.3
8:00 210.9 F 4.2
10:00 225.6 F 9.0
12:00 234.4 F 10.7
14:00 240.3 F 11.9
16:00 239.9 F 10.8
18:00 253.8 F 17.8
20:00 265.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.8 F 29.5
2:00 272.3 F 29.6
3:00 272.6 F 29.3
4:00 272.3 F 29.6

Min 271.8 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 188.7 F 0.0
10:00 206.3 F 0.0
15:00 215.1 F 0.0
20:00 213.9 F 0.0
25:00 211.5 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

01/30/2025 06:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____