

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000350  
8:53 20/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:13  
STERILIZING 3:11  
136.0 C 220kPa 3:11  
136.4 C 233kPa 3:41  
136.4 C 233kPa 4:11  
136.3 C 232kPa 4:41  
135.9 C 229kPa 5:11  
136.4 C 232kPa 5:41  
136.3 C 231kPa 6:11  
136.2 C 233kPa 6:41  
136.2 C 233kPa 6:41  
VENTING 6:41  
AIR DRYING 7:33  
CYCLE COMPLETE 7:49  
-----

Digital Signature #  
FB5462B27159F2A7  
-----