

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002071  
11:13 20/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:11  
STERILIZING 3:16  
136.0 C 220kPa 3:16  
136.3 C 225kPa 3:46  
136.4 C 227kPa 4:16  
136.3 C 225kPa 4:46  
136.3 C 226kPa 5:16  
136.1 C 222kPa 5:46  
136.4 C 226kPa 6:16  
136.3 C 224kPa 6:46  
136.3 C 224kPa 6:46  
VENTING 6:46  
AIR DRYING 7:35  
CYCLE COMPLETE 16:06  
-----

Digital Signature #  
462CBFDB6182BB99  
-----