
Midmark M11 - v1.0.5
Total Cycles: 4504

Name: RITTER2
Sterilizer ID: V2313228
04/10/2025 04:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.4 F 0.0
2:00 128.1 F 1.5
4:00 143.4 F 4.7
6:00 173.7 F 9.1
8:00 198.0 F 8.4
10:00 221.5 F 6.3
12:00 233.9 F 10.7
14:00 230.7 F 7.5
16:00 242.6 F 12.3
18:00 239.4 F 10.3
20:00 252.0 F 16.5
22:00 262.9 F 23.0
24:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.3 F 29.5
2:00 271.2 F 28.5
3:00 272.7 F 30.0
4:00 272.0 F 29.1

Min 271.0 F 28.5
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 234.5 F 0.0
10:00 219.2 F 0.0
15:00 186.6 F 0.0
20:00 177.8 F 0.0
25:00 174.7 F 0.0
30:00 181.3 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:47
CYCLE COMPLETE

04/10/2025 05:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____