

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4046

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2023 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 184.7 F 2.1  
4:00 190.5 F 6.6  
6:00 204.2 F 11.8  
8:00 217.9 F 5.8  
10:00 230.1 F 10.9  
12:00 231.0 F 7.4  
14:00 244.7 F 13.5  
16:00 243.6 F 12.0  
18:00 256.2 F 18.5  
20:00 267.5 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.1  
1:00 272.6 F 28.9  
2:00 271.7 F 28.1  
3:00 272.7 F 28.9  
4:00 271.8 F 28.2  
  
Min 271.4 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 228.5 F 0.0  
10:00 228.0 F 0.0  
15:00 230.8 F 0.0  
20:00 230.3 F 0.0  
25:00 230.8 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 2:21  
HEATING: 21:15  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

11/21/2023 11:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_