
Midmark M11 - v1.0.5
Total Cycles: 906

Name: RITTER2
Sterilizer ID: V2313228
03/25/2022 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 165.1 F 0.2
4:00 186.6 F 1.8
6:00 211.4 F 6.5
8:00 225.5 F 11.7
10:00 228.5 F 7.7
12:00 241.6 F 13.4
14:00 241.3 F 12.0
16:00 241.2 F 11.3
18:00 254.7 F 18.3
20:00 266.4 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.7
2:00 271.9 F 29.0
3:00 271.8 F 29.5
4:00 272.5 F 29.5

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 217.7 F 0.0
10:00 233.0 F 0.0
15:00 228.0 F 0.0
20:00 231.8 F 0.0
25:00 227.5 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

03/25/2022 10:55 AM
ACCEPTED: X
REJECTED:
Operator: U02