
Midmark M11 - v1.0.0
Total Cycles: 4982

Name: RITTER 1
Sterilizer ID: V2178198
10/09/2024 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.4 F 0.0
2:00 94.3 F 0.0
4:00 113.6 F 0.4
6:00 139.4 F 1.0
8:00 168.3 F 2.4
10:00 188.7 F 5.3
12:00 211.0 F 9.1
14:00 220.3 F 12.8
16:00 221.9 F 6.7
18:00 233.2 F 11.2
20:00 231.2 F 7.5
22:00 244.4 F 13.2
24:00 242.2 F 11.5
26:00 254.1 F 17.5
28:00 264.5 F 23.7
30:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.5 F 28.7
2:00 272.4 F 29.1
3:00 271.8 F 28.6
4:00 271.2 F 28.5

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 200.4 F 0.0
10:00 191.6 F 0.0
15:00 187.4 F 0.0
20:00 195.8 F 0.0
25:00 207.8 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:08:39
CYCLE COMPLETE

10/09/2024 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____