

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005558  
8:15 17/05/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:35  
STERILIZING 3:33  
136.4 C 220kPa 3:33  
136.3 C 219kPa 4:03  
136.4 C 221kPa 4:33  
136.4 C 223kPa 5:03  
136.2 C 220kPa 5:33  
136.3 C 220kPa 6:03  
136.4 C 224kPa 6:33  
136.4 C 223kPa 7:03  
136.4 C 223kPa 7:03  
VENTING 7:03  
AIR DRYING 7:51  
CYCLE COMPLETE 12:53

-----

Digital Signature #  
CF4E1BE89977B7B5  
-----