
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006422
8:06 10/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:49
136.4 C 221kPa 4:49
136.3 C 222kPa 5:19
136.3 C 222kPa 5:49
136.4 C 222kPa 6:19
136.2 C 217kPa 6:49
136.4 C 221kPa 7:19
136.3 C 219kPa 7:49
136.0 C 220kPa 8:19
136.0 C 220kPa 8:19
VENTING 8:19
AIR DRYING 9:19
CYCLE COMPLETE 9:50

Digital Signature #
738A87CAC55A9A76
