
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004285
8:37 02/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:13
136.4 C 222kPa 4:13
135.9 C 220kPa 4:43
136.4 C 221kPa 5:13
136.4 C 218kPa 5:43
136.4 C 222kPa 6:13
136.3 C 221kPa 6:43
136.3 C 220kPa 7:13
136.4 C 222kPa 7:43
136.4 C 222kPa 7:43
VENTING 7:43
AIR DRYING 8:30
CYCLE COMPLETE 12:25

Digital Signature #
2FD44FAAEA855FFA
