
Midmark M11 - v1.0.5
Total Cycles: 2101

Name: RITTER2
Sterilizer ID: V2313228
02/28/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 174.5 F 3.8
4:00 184.4 F 8.2
6:00 210.0 F 11.4
8:00 227.2 F 8.9
10:00 235.2 F 9.9
12:00 242.4 F 12.7
14:00 242.0 F 11.9
16:00 255.5 F 18.9
18:00 267.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.0 F 29.5
2:00 272.2 F 29.5
3:00 272.4 F 29.8
4:00 272.5 F 30.0

Min 271.8 F 29.5
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 233.3 F 0.0
10:00 228.1 F 0.0
15:00 223.3 F 0.0
20:00 224.8 F 0.0
25:00 227.6 F 0.0
30:00 224.3 F 0.0

Drying Complete

FILLING: 1:25
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

02/28/2023 02:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____