
Midmark M11 - v1.0.5
Total Cycles: 3090

Name: RITTER2
Sterilizer ID: V2313228
11/06/2023 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 158.1 F 2.3
4:00 178.2 F 5.8
6:00 197.1 F 10.1
8:00 215.8 F 6.7
10:00 229.5 F 8.1
12:00 240.1 F 13.0
14:00 238.2 F 10.4
16:00 235.9 F 8.7
18:00 250.2 F 15.6
20:00 261.6 F 22.2
22:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.1 F 28.7
3:00 272.4 F 29.5
4:00 271.6 F 28.9

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 228.3 F 0.0
10:00 199.7 F 0.0
15:00 191.3 F 0.0
20:00 188.2 F 0.0
25:00 186.1 F 0.0
30:00 185.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

11/06/2023 05:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____