
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003522
8:09 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:06
STERILIZING 4:13
136.0 C 220kPa 4:13
136.4 C 222kPa 4:43
136.4 C 224kPa 5:13
136.4 C 223kPa 5:43
136.1 C 222kPa 6:13
136.3 C 222kPa 6:43
136.4 C 223kPa 7:13
136.4 C 222kPa 7:43
136.4 C 222kPa 7:43
VENTING 7:43
AIR DRYING 8:35
CYCLE COMPLETE 8:40

Digital Signature #
D5443720BC77F663
