
Midmark M11 - v1.0.0
Total Cycles: 3406

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 190.6 F 3.2
4:00 190.0 F 7.5
6:00 229.4 F 13.0
8:00 233.8 F 7.8
10:00 245.8 F 13.1
12:00 241.0 F 10.5
14:00 236.0 F 8.6
16:00 251.3 F 15.5
18:00 263.0 F 22.3
20:00 272.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.6 F 28.4
2:00 271.2 F 27.8
3:00 272.7 F 28.6
4:00 271.5 F 27.9

Min 271.2 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 223.8 F 0.0
10:00 235.4 F 0.0
15:00 232.4 F 0.0
20:00 225.5 F 0.0
25:00 225.8 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

06/19/2023 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____