
Midmark M11 - v1.0.5
Total Cycles: 2018

Name: RITTER2
Sterilizer ID: V2313228
02/07/2023 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 147.7 F 0.0
4:00 169.9 F 0.7
6:00 200.5 F 4.7
8:00 215.9 F 9.2
10:00 226.0 F 9.1
12:00 234.0 F 10.4
14:00 234.0 F 8.1
16:00 241.0 F 11.0
18:00 247.0 F 14.1
20:00 259.5 F 21.0
22:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.6 F 29.2
2:00 272.8 F 29.6
3:00 272.0 F 29.1
4:00 271.4 F 28.7

Min 271.2 F 28.6
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 219.2 F 0.0
10:00 232.7 F 0.0
15:00 233.3 F 0.0
20:00 229.2 F 0.0
25:00 230.7 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:42
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

02/07/2023 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____