
Midmark M11 - v1.0.5
Total Cycles: 23

Name: RITTER2
Sterilizer ID: V2313228
04/27/2021 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.9 F 0.0
2:00 148.4 F 2.0
4:00 153.6 F 5.3
6:00 179.2 F 9.7
8:00 205.1 F 5.7
10:00 218.5 F 8.7
12:00 232.0 F 11.0
14:00 238.1 F 11.9
16:00 239.7 F 10.7
18:00 252.8 F 17.5
20:00 264.7 F 24.5
22:00 270.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.7
1:00 272.7 F 30.0
2:00 272.7 F 30.5
3:00 272.1 F 29.6
4:00 272.8 F 30.3

Min 271.8 F 29.6
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 222.9 F 0.0
10:00 235.9 F 0.0
15:00 215.8 F 0.0
20:00 210.0 F 0.0
25:00 205.8 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

04/27/2021 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____