
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004767
10:23 28/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:11
STERILIZING 4:15
136.0 C 220kPa 4:15
136.2 C 221kPa 4:45
136.1 C 221kPa 5:15
136.1 C 222kPa 5:45
136.1 C 222kPa 6:15
136.4 C 223kPa 6:45
136.3 C 221kPa 7:15
136.3 C 222kPa 7:45
136.3 C 222kPa 7:45
VENTING 7:45
AIR DRYING 8:36
CYCLE COMPLETE 8:38

Digital Signature #
581E351CE4651A2F
