
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000799
8:57 17/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:32
PRESSURIZING 7:39
STERILIZING 9:34
136.4 C 222kPa 9:34
136.1 C 221kPa 10:04
136.3 C 223kPa 10:34
136.4 C 224kPa 11:04
136.4 C 222kPa 11:34
136.4 C 223kPa 12:04
135.8 C 218kPa 12:34
136.3 C 223kPa 13:04
136.3 C 223kPa 13:04
VENTING 13:04
AIR DRYING 14:08
CYCLE COMPLETE 39:36

Digital Signature #
0E69D9250C40AC44
