
Midmark M11 - v1.0.0
Total Cycles: 1611

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.0 F 0.0
2:00 166.5 F 3.5
4:00 182.3 F 8.1
6:00 209.7 F 10.5
8:00 221.0 F 7.8
10:00 235.9 F 13.0
12:00 237.0 F 10.1
14:00 236.4 F 8.4
16:00 249.4 F 15.0
18:00 261.4 F 21.8
20:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.3 F 28.8
2:00 271.6 F 28.5
3:00 272.6 F 28.9
4:00 271.5 F 28.2

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 228.3 F 0.0
10:00 229.7 F 0.0
15:00 230.0 F 0.0
20:00 230.3 F 0.0
25:00 228.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

02/08/2022 05:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____