
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001566
12:12 11/10/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:01
PRESSURIZING 7:02
STERILIZING 8:45
136.0 C 220kPa 8:45
136.0 C 218kPa 9:15
136.2 C 218kPa 9:45
136.4 C 222kPa 10:15
136.4 C 222kPa 10:45
135.9 C 219kPa 11:15
136.5 C 221kPa 11:45
135.9 C 220kPa 12:15
VENTING 12:15
AIR DRYING 13:22
CYCLE COMPLETE 22:38

Digital Signature #
524D0C981AA82A38
