
Midmark M11 - v1.0.5
Total Cycles: 804

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.2 F 0.0
2:00 159.9 F 3.0
4:00 175.8 F 7.0
6:00 200.1 F 12.3
8:00 218.0 F 6.8
10:00 232.9 F 12.4
12:00 236.4 F 10.1
14:00 236.0 F 8.5
16:00 250.5 F 16.0
18:00 263.0 F 23.1
20:00 271.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.7 F 29.6
2:00 272.0 F 29.5
3:00 272.7 F 29.6
4:00 271.9 F 29.3

Min 271.8 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 220.2 F 0.0
10:00 233.2 F 0.0
15:00 223.8 F 0.0
20:00 220.6 F 0.0
25:00 219.9 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

02/15/2022 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____