
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002916
9:56 10/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:49
STERILIZING 4:45
136.0 C 220kPa 4:45
136.3 C 222kPa 5:15
136.4 C 224kPa 5:45
136.4 C 224kPa 6:15
136.1 C 221kPa 6:45
136.4 C 223kPa 7:15
136.2 C 217kPa 7:45
136.2 C 223kPa 8:15
136.2 C 223kPa 8:15
VENTING 8:15
AIR DRYING 9:09
CYCLE COMPLETE 9:11

Digital Signature #
F8C65E0C0733224E
