

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002595  
10:41 24/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:48  
STERILIZING 3:51  
136.5 C 224kPa 3:51  
136.3 C 225kPa 4:21  
136.4 C 223kPa 4:51  
136.3 C 222kPa 5:21  
136.2 C 224kPa 5:51  
136.1 C 222kPa 6:21  
136.2 C 223kPa 6:51  
136.4 C 224kPa 7:21  
136.4 C 224kPa 7:21  
VENTING 7:21  
AIR DRYING 8:13  
CYCLE COMPLETE 25:58  
-----

Digital Signature #  
657BF03FFE2CE0A5  
-----