
Midmark M11 - v1.0.5
Total Cycles: 3485

Name: RITTER2
Sterilizer ID: V2313228
03/27/2024 05:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.7 F 0.0
2:00 174.1 F 3.7
4:00 201.3 F 9.8
6:00 220.7 F 5.2
8:00 236.1 F 12.1
10:00 239.2 F 10.6
12:00 238.5 F 10.0
14:00 254.1 F 17.8
16:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.4 F 29.1
2:00 272.6 F 29.6
3:00 272.1 F 28.7
4:00 273.0 F 29.6

Min 271.6 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 204.0 F 0.0
10:00 204.9 F 0.0
15:00 205.6 F 0.0
20:00 204.0 F 0.0
25:00 201.7 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 17:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:56:24
CYCLE COMPLETE

03/27/2024 06:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____