
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000060
8:28 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:09
STERILIZING 4:33
136.0 C 220kPa 4:33
136.2 C 221kPa 5:03
136.4 C 224kPa 5:33
136.4 C 225kPa 6:03
136.4 C 224kPa 6:33
136.3 C 224kPa 7:03
136.4 C 225kPa 7:33
135.9 C 220kPa 8:03
135.9 C 220kPa 8:03
VENTING 8:03
AIR DRYING 8:57
CYCLE COMPLETE 23:57

Digital Signature #
D7336C808E409D9A
