

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002575  
9:30 14/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:30  
STERILIZING 3:31  
136.0 C 220kPa 3:31  
136.1 C 222kPa 4:01  
136.5 C 225kPa 4:31  
136.4 C 225kPa 5:01  
136.4 C 221kPa 5:31  
136.4 C 222kPa 6:01  
136.1 C 223kPa 6:31  
136.0 C 220kPa 7:01  
136.0 C 220kPa 7:01  
VENTING 7:01  
AIR DRYING 7:52  
CYCLE COMPLETE 37:50  
-----

Digital Signature #  
EA2A05268F8ED92B  
-----