
Midmark M11 - v1.0.5
Total Cycles: 4045

Name: RITTER2
Sterilizer ID: V2313228
10/22/2024 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 89.1 F 0.0
4:00 83.7 F 0.0
6:00 105.6 F 0.1
8:00 145.1 F 0.2
10:00 164.9 F 0.6
12:00 190.5 F 3.8
14:00 224.0 F 7.0
16:00 232.5 F 10.7
18:00 227.4 F 7.8
20:00 234.3 F 9.5
22:00 237.9 F 10.9
24:00 241.7 F 11.6
26:00 238.8 F 10.0
28:00 251.5 F 16.2
30:00 262.2 F 22.4
32:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 29.1
2:00 271.7 F 28.8
3:00 272.3 F 29.1
4:00 271.6 F 28.7

Min 271.5 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 211.0 F 0.0
10:00 192.9 F 0.0
15:00 188.3 F 0.0
20:00 183.3 F 0.0
25:00 198.9 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:30
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:43
CYCLE COMPLETE

10/22/2024 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____