
Midmark M11 - v1.0.0
Total Cycles: 3106

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 189.3 F 3.8
4:00 207.5 F 8.8
6:00 223.0 F 9.3
8:00 231.2 F 9.5
10:00 236.3 F 9.5
12:00 243.4 F 12.6
14:00 241.9 F 11.4
16:00 254.9 F 18.0
18:00 266.0 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.9
2:00 271.9 F 28.8
3:00 272.3 F 28.8
4:00 271.8 F 28.8

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 227.6 F 0.0
10:00 236.1 F 0.0
15:00 229.4 F 0.0
20:00 222.3 F 0.0
25:00 221.0 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

04/13/2023 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____