

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5385

Name: RITTER 1  
Sterilizer ID: V2178198  
03/04/2025 02:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 84.7 F 0.0  
2:00 96.3 F 0.6  
4:00 99.0 F 2.6  
6:00 126.0 F 5.4  
8:00 160.0 F 8.6  
10:00 175.5 F 12.2  
12:00 205.2 F 3.2  
14:00 214.4 F 6.4  
16:00 222.5 F 10.0  
18:00 234.8 F 10.1  
20:00 237.0 F 10.1  
22:00 236.4 F 8.9  
24:00 245.6 F 13.3  
26:00 256.1 F 19.0  
28:00 266.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.4 F 29.6  
2:00 272.4 F 29.1  
3:00 271.2 F 28.7  
4:00 271.8 F 29.4

Min 271.2 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 196.3 F 0.0  
10:00 204.0 F 0.0  
15:00 207.1 F 0.0  
20:00 212.1 F 0.0  
25:00 216.3 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 29:52  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:08:32  
CYCLE COMPLETE

03/04/2025 03:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_