
Midmark M11 - v1.0.0
Total Cycles: 3040

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.0 F 0.0
2:00 202.7 F 4.8
4:00 213.7 F 10.9
6:00 226.2 F 6.8
8:00 238.5 F 12.7
10:00 240.1 F 10.8
12:00 238.0 F 9.5
14:00 253.1 F 16.8
16:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.4 F 28.8
2:00 271.5 F 28.3
3:00 272.1 F 28.6
4:00 271.6 F 28.5

Min 271.4 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 228.8 F 0.0
10:00 237.3 F 0.0
15:00 241.8 F 0.0
20:00 226.8 F 0.0
25:00 225.1 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 17:35
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

03/27/2023 04:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____