
Midmark M11 - v1.0.0
Total Cycles: 4085

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.6 F 0.0
2:00 97.0 F 0.8
4:00 109.8 F 2.7
6:00 136.9 F 5.3
8:00 167.1 F 8.4
10:00 195.5 F 12.4
12:00 205.3 F 3.5
14:00 214.3 F 6.8
16:00 225.1 F 10.6
18:00 230.9 F 7.6
20:00 238.0 F 11.0
22:00 235.4 F 8.5
24:00 248.6 F 14.6
26:00 259.5 F 20.6
28:00 269.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.7 F 29.2
2:00 272.0 F 28.7
3:00 271.2 F 28.3
4:00 272.3 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 214.0 F 0.0
10:00 228.1 F 0.0
15:00 234.8 F 0.0
20:00 227.3 F 0.0
25:00 228.1 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:03
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:07:42
CYCLE COMPLETE

12/05/2023 03:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____