
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000983
11:42 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:17
STERILIZING 3:20
136.4 C 232kPa 3:20
136.0 C 230kPa 3:50
136.2 C 232kPa 4:20
136.1 C 233kPa 4:50
136.3 C 232kPa 5:20
136.3 C 232kPa 5:50
136.4 C 231kPa 6:20
136.3 C 231kPa 6:50
136.3 C 231kPa 6:50
VENTING 6:50
AIR DRYING 7:40
CYCLE COMPLETE 7:48

Digital Signature #
F47B97123AD968D3
