
Midmark M11 - v1.0.5
Total Cycles: 2768

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 164.1 F 0.0
4:00 188.9 F 0.0
6:00 209.1 F 2.7
8:00 223.8 F 7.3
10:00 236.4 F 12.2
12:00 236.2 F 9.8
14:00 237.5 F 9.2
16:00 240.9 F 10.6
18:00 248.8 F 14.9
20:00 261.1 F 21.8
22:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.0 F 28.8
2:00 272.6 F 29.2
3:00 272.2 F 29.3
4:00 271.7 F 28.6

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 214.7 F 0.0
10:00 212.7 F 0.0
15:00 207.4 F 0.0
20:00 203.8 F 0.0
25:00 208.2 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

08/15/2023 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____