

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1783

Name: RITTER 1  
Sterilizer ID: V2178198  
04/01/2022 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
2:00 187.6 F 3.7  
4:00 208.2 F 9.9  
6:00 220.1 F 5.0  
8:00 237.5 F 11.2  
10:00 236.7 F 9.4  
12:00 236.4 F 8.6  
14:00 250.9 F 15.8  
16:00 263.5 F 23.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 271.7 F 28.4  
2:00 272.0 F 28.6  
3:00 272.5 F 28.9  
4:00 272.4 F 29.2  
  
Min 271.2 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
5:00 231.6 F 0.0  
10:00 231.1 F 0.0  
15:00 229.0 F 0.0  
20:00 230.7 F 0.0  
25:00 229.8 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 17:57  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:56:58  
CYCLE COMPLETE

04/01/2022 12:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_