
Midmark M11 - v1.0.5
Total Cycles: 587

Name: RITTER2
Sterilizer ID: V2313228
11/30/2021 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.0 F 0.0
2:00 149.9 F 2.4
4:00 167.3 F 6.1
6:00 190.8 F 11.0
8:00 212.7 F 4.9
10:00 227.2 F 9.9
12:00 232.9 F 9.0
14:00 241.6 F 12.5
16:00 240.5 F 11.1
18:00 252.8 F 17.3
20:00 263.5 F 23.6
22:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.9 F 29.8
2:00 271.8 F 28.9
3:00 272.8 F 29.9
4:00 272.4 F 29.3

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 228.6 F 0.0
10:00 220.2 F 0.0
15:00 222.5 F 0.0
20:00 222.4 F 0.0
25:00 219.3 F 0.0
30:00 208.9 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

11/30/2021 12:04 PM
ACCEPTED: X
REJECTED:
Operator: U02