
Midmark M11 - v1.0.5
Total Cycles: 144

Name: RITTER2
Sterilizer ID: V2313228
06/25/2021 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.5 F 0.0
2:00 122.2 F 1.4
4:00 141.3 F 4.2
6:00 165.1 F 8.6
8:00 197.4 F 10.4
10:00 210.1 F 7.9
12:00 226.1 F 13.3
14:00 231.6 F 10.5
16:00 233.7 F 8.7
18:00 246.4 F 15.4
20:00 258.3 F 21.9
22:00 269.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 30.3
1:00 272.1 F 30.2
2:00 272.3 F 30.3
3:00 272.4 F 30.5
4:00 272.6 F 30.8

Min 272.0 F 30.1
Max 273.3 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 228.1 F 0.0
10:00 211.9 F 0.0
15:00 204.6 F 0.0
20:00 212.9 F 0.0
25:00 228.8 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:51
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

06/25/2021 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____