
Midmark M11 - v1.0.0
Total Cycles: 4620

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2024 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 70.2 F 0.0
4:00 76.4 F 0.2
6:00 111.5 F 0.5
8:00 143.8 F 2.1
10:00 167.5 F 5.1
12:00 186.6 F 8.3
14:00 199.3 F 11.4
16:00 214.5 F 4.7
18:00 219.3 F 6.9
20:00 229.0 F 10.5
22:00 232.7 F 8.4
24:00 239.0 F 11.1
26:00 235.0 F 8.5
28:00 248.6 F 14.7
30:00 259.4 F 20.6
32:00 269.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.1 F 29.1
4:00 272.3 F 28.8

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 198.4 F 0.0
10:00 186.9 F 0.0
15:00 182.6 F 0.0
20:00 182.8 F 0.0
25:00 183.5 F 0.0
30:00 189.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:06
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:30
CYCLE COMPLETE

06/10/2024 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____