
Midmark M11 - v1.0.5
Total Cycles: 1297

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 170.8 F 0.3
4:00 196.3 F 4.6
6:00 214.0 F 10.0
8:00 229.3 F 7.1
10:00 235.5 F 11.1
12:00 235.2 F 8.9
14:00 240.3 F 10.1
16:00 248.1 F 14.5
18:00 260.2 F 21.3
20:00 271.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.1
2:00 271.5 F 29.3
3:00 272.2 F 28.9
4:00 272.2 F 29.9

Min 271.0 F 28.6
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 230.3 F 0.0
10:00 223.7 F 0.0
15:00 188.4 F 0.0
20:00 176.7 F 0.0
25:00 191.3 F 0.0
30:00 174.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

07/14/2022 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____