
Midmark M11 - v1.0.5
Total Cycles: 865

Name: RITTER2
Sterilizer ID: V2313228
03/09/2022 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 196.8 F 3.9
4:00 201.3 F 9.3
6:00 216.8 F 7.7
8:00 230.9 F 10.7
10:00 232.2 F 8.3
12:00 239.6 F 10.5
14:00 246.9 F 14.4
16:00 259.5 F 21.4
18:00 270.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.5 F 29.9
2:00 271.7 F 29.2
3:00 271.8 F 29.8
4:00 272.6 F 29.8

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 197.1 F 0.0
10:00 209.2 F 0.0
15:00 214.3 F 0.0
20:00 216.0 F 0.0
25:00 216.2 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

03/09/2022 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____