
Midmark M11 - v1.0.0
Total Cycles: 4743

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2024 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 70.4 F 0.0
4:00 84.7 F 0.1
6:00 115.6 F 0.6
8:00 146.2 F 3.0
10:00 170.1 F 6.4
12:00 185.5 F 10.1
14:00 201.6 F 8.2
16:00 214.3 F 5.7
18:00 225.8 F 9.5
20:00 235.4 F 12.1
22:00 236.8 F 10.0
24:00 236.5 F 8.7
26:00 247.0 F 13.7
28:00 258.6 F 20.0
30:00 269.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 196.0 F 0.0
10:00 188.5 F 0.0
15:00 190.8 F 0.0
20:00 206.1 F 0.0
25:00 214.2 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

07/18/2024 09:16 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____