
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001925
9:57 05/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:16
STERILIZING 4:21
136.0 C 220kPa 4:21
136.3 C 220kPa 4:51
136.3 C 219kPa 5:21
136.3 C 224kPa 5:51
136.4 C 224kPa 6:21
136.4 C 224kPa 6:51
136.1 C 222kPa 7:21
136.2 C 223kPa 7:51
136.2 C 223kPa 7:51
VENTING 7:51
AIR DRYING 8:45
CYCLE COMPLETE 9:02

Digital Signature #
F1E87CAD6F1C1FB1
