

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001489  
12:38 26/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:47  
STERILIZING 3:47  
136.0 C 221kPa 3:47  
136.4 C 226kPa 4:17  
136.2 C 224kPa 4:47  
136.4 C 226kPa 5:17  
136.4 C 226kPa 5:47  
136.4 C 224kPa 6:17  
136.4 C 223kPa 6:47  
136.4 C 226kPa 7:17  
136.4 C 226kPa 7:17  
VENTING 7:17  
AIR DRYING 8:12  
CYCLE COMPLETE 8:16  
-----

Digital Signature #  
4641A981D63B2742  
-----