
Midmark M11 - v1.0.0
Total Cycles: 4362

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2024 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 79.8 F 0.2
4:00 103.7 F 1.0
6:00 114.1 F 3.3
8:00 147.2 F 5.9
10:00 167.2 F 8.9
12:00 185.8 F 12.5
14:00 206.6 F 2.9
16:00 215.8 F 5.9
18:00 224.5 F 9.0
20:00 233.6 F 12.7
22:00 233.1 F 8.0
24:00 243.9 F 12.8
26:00 240.7 F 10.4
28:00 252.3 F 16.2
30:00 262.4 F 22.1
32:00 272.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.9
4:00 272.4 F 28.4

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 228.8 F 0.0
10:00 218.5 F 0.0
15:00 218.8 F 0.0
20:00 213.8 F 0.0
25:00 215.2 F 0.0
30:00 216.7 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:34
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:56
CYCLE COMPLETE

03/07/2024 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____