
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002275
9:31 18/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:48
136.0 C 220kPa 3:48
135.9 C 220kPa 4:18
136.3 C 224kPa 4:48
136.4 C 225kPa 5:18
136.1 C 224kPa 5:48
136.3 C 223kPa 6:18
136.2 C 223kPa 6:48
136.2 C 222kPa 7:18
136.2 C 222kPa 7:18
VENTING 7:19
AIR DRYING 8:11
CYCLE COMPLETE 32:53

Digital Signature #
1A19D6CBC2DD1D25
