
Midmark M11 - v1.0.5
Total Cycles: 2504

Name: RITTER2
Sterilizer ID: V2313228
06/13/2023 07:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 78.3 F 0.0
4:00 88.7 F 0.0
6:00 101.0 F 0.0
8:00 148.6 F 0.4
10:00 171.6 F 3.6
12:00 189.0 F 7.1
14:00 208.9 F 10.9
16:00 219.0 F 6.2
18:00 228.3 F 9.0
20:00 239.3 F 13.5
22:00 238.3 F 10.9
24:00 236.1 F 9.1
26:00 249.5 F 15.5
28:00 260.7 F 21.9
30:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 28.8
3:00 272.5 F 29.8
4:00 272.2 F 29.2

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 218.7 F 0.0
10:00 229.1 F 0.0
15:00 232.5 F 0.0
20:00 227.2 F 0.0
25:00 230.7 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

06/13/2023 08:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____