
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000533
15:24 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:21
STERILIZING 3:22
136.0 C 227kPa 3:22
136.1 C 231kPa 3:52
136.4 C 233kPa 4:22
136.3 C 232kPa 4:52
136.1 C 229kPa 5:22
136.4 C 232kPa 5:52
136.1 C 230kPa 6:22
136.4 C 233kPa 6:52
136.4 C 233kPa 6:52
VENTING 6:52
AIR DRYING 7:44
CYCLE COMPLETE 22:44

Digital Signature #
9232607EE8CF0C16
