
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006662
14:51 22/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:25
PRESSURIZING 1:33
STERILIZING 2:27
136.0 C 220kPa 2:27
136.4 C 222kPa 2:57
136.5 C 225kPa 3:27
136.4 C 224kPa 3:57
136.2 C 223kPa 4:27
136.4 C 225kPa 4:57
136.4 C 221kPa 5:27
136.4 C 225kPa 5:57
136.4 C 225kPa 5:57
VENTING 5:57
AIR DRYING 6:48
CYCLE COMPLETE 6:53

Digital Signature #
D6032A37D811B051
